


Difference Spectrum and Model Fit



meas_MID00199_FID32114_CBU...
Filename: _svs_mpress_GABA_INS.nii.gz

Area

GABA+: 0.000524
Glx: 0.000525
Water: 10
Cr: 0.0055

Linewidth

Water: 8.42 Hz
Cr: 8.67 Hz

SNR

Water: 208456
Cr: 286

Fit Error

GABA+,Water: 5.25%
GABA+,Cr: 5.35%
Glx,Water: 3.06%
Glx,Cr: 3.22%

Quantification

GABA+/Water: 1.87 i.u.
GABA+/Cr: 0.10
Glx/Water: 6.90 i.u.
Glx/Cr: 0.10
FitVer: 230410

Reference Signals

